Piezoelectric Accelerometer

Model 2228C

- Triaxial
- Ground Isolated
- Light Weight (15 gm)
- Requires No External Power
- Vibration Measurement in Three Orthogonal Axis



DESCRIPTION

The ENDEVCO[®] Model 2228C is a small triaxial piezoelectric accelerometer designed specifically for vibration measurement in three orthogonal axes on small structures and objects. The transducer features three 10-32 receptacles for output connection and can be screw or adhesively mounted. Its light weight (15 gm) effectively minimizes mass loading. The accelerometer is a self-generating device that requires no external power source for operation.

The Model 2228C features ENDEVCO's PIEZITE[®] Type P-8 crystal elements, operating in annular shear mode, this unit exhibits excellent output sensitivity stability over time. Signal ground is isolated from the mounting surface of the unit. Low-noise, flexible coaxial cables are supplied for error-free operation.

ENDEVCO Signal Conditioner Models 133, 2775A or OASIS 2000 Computer-Controlled System are recommended for use with this high impedance accelerometer.

中国


:ual size


ENDEVCO MODEL 2228C


北京市朝阳区望京西路48号 金隅国际C座1002 电话:+86 010 8477 5646 传真:+86 010 5894 9029

北京赛斯维测控技术有限公司

邮箱:sales@sensorway.cn













ENDEVCO MODEL 2228C

Piezoelectric Accelerometer

SPECIFICATIONS

The following performance specifications conform to ISA-RP-37.2 (1964) and are typical values, referenced at +75°F (+24°C) and 100 Hz, unless otherwise noted. Calibration data, traceable to National Institute of Standards and Technology (NIST), is supplied.

DYNAMIC CHARACTERISTICS		Units	
CHARGE SENSITIVITY			
TYPICAL		pC/g	2.8
MINIMUM		pC/g	2.2
FREQUENCY RESPONSE			See Typical Amplitude Response
RESONANCE FREQUENCY		kHz	21 (Typical)
AMPLITUDE RESPONS	E [1]		
±5%		Hz	1 to 4000
±1dB		Hz	.5 to 4000 (x,y) .5 to 5000 (z)
TEMPERATURE RESPO	TEMPERATURE RESPONSE		See Typical Curve
TRANSVERSE SENSITIVITY		%	≤ 5
AMPLITUDE LINEARITY [2]		%	1
Per 500 g, 0 to 2000 g			
ELECTRICAL CHARACTERISTICS			
OUTPUT POLARITY			Acceleration applied in the direction of the arrow
Con on Cerain			the unit produces positive output
RESISTANCE		GΩ	≥ 10
ISOLATION		MΩ	≥ 10
	anal around and to mounting		≥ 10
Signal ground to each signal ground and to mounting surface CAPACITANCE DF 400			
GROUNDING		pF	Each sensor is isolated from the anodized
GROUNDING			
			aluminum case
ENVIRONMENTAL CHARACTERISTICS			
TEMPERATURE RANGE			-67°F to +350°F (-55°C to +177°C)
HUMIDITY			Epoxy sealed, non-hermetic
SINUSOIDAL VIBRATIO	N LIMIT	g pk	1000
SHOCK LIMIT		g pk	2000
ELECTROMAGNETIC S	ENSITIVITY	equiv. g rms/gauss	0.01
PHYSICAL CHARACTE DIMENSIONS	RISTICS		See Outline Drawing
WEIGHT		gm (oz)	15 (0.53)
CASE MATERIAL			Aluminum alloy case, nickel alloy sensors
CONNECTOR			Mates with Endevco 3060 Series Cable
MOUNTING/TORQUE		lbf-in (Nm)	8 (1)
CALIBRATION			
SUPPLIED:			
CHARGE SENSITIVITY		pC/g	
CAPACITANCE		pF	
MAXIMUM TRANSVERS		%	
CHARGE FREQUENCY	RESPONSE	%	20 to 4000 Hz
		dB	thru resonance (Z axis only)
	CABLE ASSEMBLY, Three Each [3] 4-40 X 5/8 SOCKET HEAD CAP SCREWS, TWO EACH NO. 4 FLAT WASHER, TWO EACH HEX WRENCH NES IN-LINE CHARGE CONVERTOR FOR USE WITH CONSTANT CURRENT SOURCE		 Adhesives such as petro-wax, hot-melt glue, and cyanoacrylate epoxy (super glue) may be used to mount the accelerometer temporarily to the test structure. An adhesive mounting kit (P/N 31849) is available as an option from Endevco. To remove an epoxy-mounted accelerometer, first soften the epoxy with an appropriate solvent and then twist the unit off with the supplied emoval wrench. Damage to sensors caused by inappropriate removal procedures are not covered by Endevco's warranty. Maintain high levels of precision and accuracy using Endevco's
 NOTES 1. Low-end response of the transducer is a function of its associated electronics. 2. Short duration shock pulses, such as those generated by metal-to-metal impacts, may excite transducer resonance and cause linearity errors. Send for TP290 for more details. 3. Flexible cable, such as the supplied 3060A, should be used to minimize cable-strain errors. 			factory calibration services. Call Endevco's inside sales force at 800-982-6732 for recommended intervals, pricing and turn- around time for these services as well as for quotations on our standard products.

Continued product improvement necessitates that Endevco reserve the right to modify these specifications without notice. Endevco maintains a program of con-stant surveillance over all products to ensure a high level of reliability. This program includes attention to reliability factors during product design, the support of stringent Quality Control requirements, and compulsory corrective action procedures. These measures, together with conservative specifications have made the name Endevco synonymous with reliability.

ENDEVCO CORPORATION, 30700 RANCHO VIEJO ROAD, SAN JUAN CAPISTRANO, CA 92675 USA (800) 982-6732 (949) 493-8181 fax (949) 661-7231 www.endevco.com Email:applications@endevco.com 0203 REV A